tenders@gysbi.com SC-FO-002-A Revision No.: 2 Date: 08 May 2025

Request for Quotation

PROVISION OF SERVICE - ACCURACY CHECK, CALIBRATION, TROUBLESHOOTING, REPROGRAMMING, AND REPAIR OF ALCO-SENSOR VXL AND FST BREATH ALCOHOL TESTERS

1. INTRODUCTION

Guyana Shore Base Inc. (GYSBI) invites qualified and experienced bidders to submit Quotation for the Supply and Delivery of Accuracy Check, Calibration, Troubleshooting, Reprogramming, and Repair Services for Alco-Sensor VXL and FST Breath Alcohol Testers for our facility at Houston, East Bank Demerara.

2. PRODUCT REQUIREMENTS

- Delivery Location: GYSBI Port Facility, Houston, East Bank Demerara.
- Objective: To provide accuracy check, and as needed, calibration adjustment, troubleshooting, reprogramming, and repair services for Evidential Breath Alcohol Testers

 Alco-Sensor VXL and FST Devices
- Key Consideration: See specifications below.

3. REQUIREMENTS FOR QUOTATION SUBMISSION

Interested contractors must provide the following documents:

- Local Content Certificate (if available). If no Local Content Certificate is available, Guyaneseregistered businesses can provide evidence of majority-Guyanese ownership and indication that local content certification is being or will be pursued.
- Valid Business Registration or equivalent
- Valid GRA & NIS Compliance
- Any Other Certification

4. SUBMISSION DETAILS

- Date published: July 11, 2025
- Deadline for Submission: August 11, 2025
- Submission Format: Single PDF document
- Submission Email Address: tenders@gysbi.com
- Clarifications & Queries: All inquiries can be submitted to the email address above

 (\underline{R})

 \sim

5. DISCLAIMER

This RFQ does not constitute a solicitation for bids or an offer to contract. Guyana Shore Base Inc. reserves the right to accept or reject any submission at its sole discretion and to proceed with the procurement in any manner it deems fit.

TECHNICAL SPECIFICATION

Description of Requirement	Specification	Quantity
Accuracy checks as needed, calibration adjustments, troubleshooting, reprogramming, and repair of Breath Alcohol Testers should any device fall outside of the acceptable tolerance level. Frequency of Accuracy checks: • Alco-Sensor VXL – Monthly • Alco-Sensor FST – Yearly (or otherwise as per manufacturer requirements)	 Provision of a Quality Assurance Plan/Maintenance Program Plan for Alco-Sensor VXL and FST Intoximeters, outlining: Company overview, service scope, and requirements Management system summary, along with applicable accreditations/ certifications Specifications of standard(s) to be used to perform accuracy checks and calibration adjustments. acceptable tolerance of the identified devices recommended minimum intervals for performing accuracy checks/calibration adjustments, and any other maintenance required as per manufacturer requirements. Accuracy checks and calibration processes to be traceable to international standards. All completed accuracy checks/calibration adjustments to be evidence through the provision of a certificate outlining all specifications of the service provided upon return of each device. 	3 VXL Devices 4 FST Devices

tenders@gysbi.com SC-FO-002-A Revision No.: 2 Date: 08 May 2025

Description of Requirement	Specification	Quantity
	All technicians providing services should be trained and certified.	
	Certificates for each technician who	
	will be providing services are to be provided for review and record-	
	keeping.	

Issued by:

GUYANA SHORE BASE INC. (GYSBI) July 11, 2025